

16S Microbiome NGS Assay

Library Preparation Kit

REF

9-131 [Set A]
9-132 [Set B]
9-133 [Set C]

RUO



96 Reactions (Library Preparations)

2°C  8°C

Magnetic Beads

-30°C  -15°C

Other Components

Upon arrival, please store components according to their labels!

ATTENTION!

MAJOR CHANGES IN PROTOCOL AND PCR2 PRIMER SEQUENCES
IN REV 3.0!
DETAILS SEE PAGE 16.

ViennaLab Diagnostics GmbH

Gaudenzdorfer Guertel 43-45

A-1120 Vienna, Austria

Phone: (+43-1) 8120156-0

info@viennalab.com



www.viennalab.com

Instructions for use

Library Preparation Kit for 16S Microbiome Analysis

Table of Contents

1. Introduction	3
2. Kit components	4
3. Equipment and reagents required but not supplied	5
3.1. Sample preparation	5
3.2. PCR.....	5
3.3. Sample purification	5
3.4. Library Quality Control	5
3.5. Library Quantification.....	5
3.6. Sequencing reagents.....	5
4. Laboratory protocol	6
4.1. First PCR – Amplification of target region	6
4.1.1. Reaction setup	6
4.1.2. PCR program	6
4.2. Cleanup of PCR1 products	7
4.2.1. Preparation	7
4.2.2. Clean-up procedure	7
4.2.3. (Optional) Check PCR1 product size	8
4.3. Second PCR – Indexing	9
4.3.1. Reaction setup	9
4.3.2. PCR program	9
4.4. Cleanup of PCR2 products	10
4.4.1. Preparation	10
4.4.2. Clean-up procedure	10
4.4.3. (Optional) Check PCR2 product size	11
4.5. Library quantification and pooling	11
4.5.1. Library quantification.....	11
4.5.2. Library dilution and pooling.....	11
5. Sample Sheet and sequencing recommendations	12
6. Quick Guide to sequencing data analysis	14
APPENDIX	15

1. Introduction

The kit is intended for preparing NGS libraries for the bacterial 16S rRNA gene-based characterization of the human microbiome. In the first PCR step, the highly variable V3-V4 regions are amplified with locus-specific primers. The second PCR introduces dual index sequences for the assignment of the reads to individual samples during data demultiplexing. The final amplicon structure contains all sequences required for the analysis of the library pool on Illumina platforms (see Figure 1). This kit has been validated for the Illumina MiSeq instrument.

Graphical Workflow:




Figure 1 - Graphical Workflow of the library preparation procedure

2. Kit components

[REF 9-131]

Activation code sticker for 100 analyses [attached inside to the of the kit box lid]

Magnetic Beads 13ml

Box "16S Master Mix 2x":	5x 1ml 16S Master Mix 2x
Box "PCR1 V3-V4 Primer Mix":	1x 400µl MIB1 – V3V4 Mix
Box "PCR2 Indexing Forward Primers Set A ":	8x 30µl MIB2A-F1 – MIB2A-F8 Primer
Box "PCR2 Indexing Reverse Primers Box 1/2 Set A ":	8x 30µl MIB2X-R1 – MIB2X-R8 Primer
Box "PCR2 Indexing Reverse Primers Box 2/2 Set A ":	4x 30µl MIB2X-R9 – MIBX2-R12 Primer

[REF 9-132]

Activation code sticker for 100 analyses [attached inside to the of the kit box lid]

Magnetic Beads 13ml

Box "16S Master Mix 2x":	5x 1ml 16S Master Mix 2x
Box "PCR1 V3-V4 Primer Mix":	1x 400µl MIB1 – V3V4 Mix
Box "PCR2 Indexing Forward Primers Set B ":	8x 30µl MIB2B-F9 – MIB2B-F16 Primer
Box "PCR2 Indexing Reverse Primers Box 1/2 Set B ":	8x 30µl MIB2X-R1 – MIB2X-R8 Primer
Box "PCR2 Indexing Reverse Primers Box 2/2 Set B ":	4x 30µl MIB2X-R9 – MIBX2-R12 Primer

[REF 9-133]

Activation code sticker for 100 analyses [attached inside to the of the kit box lid]

Magnetic Beads 13ml

Box "16S Master Mix 2x":	5x 1ml 16S Master Mix 2x
Box "PCR1 V3-V4 Primer Mix":	1x 400µl MIB1 – V3V4 Mix
Box "PCR2 Indexing Forward Primers Set C ":	8x 30µl MIB2C-F17 – MIB2C-F24 Primer
Box "PCR2 Indexing Reverse Primers Box 1/2 Set C ":	8x 30µl MIB2X-R1 – MIB2X-R8 Primer
Box "PCR2 Indexing Reverse Primers Box 2/2 Set C ":	4x 30µl MIB2X-R9 – MIBX2-R12 Primer

Note: REF 9-131, 9-132, 9-133 differ only in Indexing Primer Sets. If you plan to sequence more than 96 samples on one flow cell, please order different sets (e.g. Set A / REF 9-131 and Set B / REF 9-132). Make sure that all individual libraries sequenced in the same pool have a unique indexing primer combination.

3. Equipment and reagents required but not supplied

3.1. Sample preparation

- Recommended for DNA isolation: Stool DNA Isolation Kit (Norgen Biotek).
Other DNA isolation kits have not been validated and may lead to incorrect results.
- Recommended for DNA quantification: fluorometric methods, such as Qubit™.
- DNA concentration range: 0.5-5ng/μl.

3.2. PCR

- PCR tubes or suitable 96-well plates.
- PCR cycler with lid heating and specified ramp rates.
- Nuclease-free, PCR-grade water (recommended: Illumina wash buffer PW1 or another molecular biology-grade water).

3.3. Sample purification

- 80% ethanol, freshly prepared.
- Magnetic separator for 1.5 ml Eppendorf tubes, PCR strips or 96-well plates.

3.4. Library Quality Control

- Equipment and reagents to prepare 2-4% agarose gel and run the PCR products.
- Alternatively, Fragment Analyzer using DNF-473, DNF-474 or DNF-477 kit.

3.5. Library Quantification

- Library quantification by qPCR is recommended, e.g. with JetSeq™ Library Quantification Lo-ROX Kit (meridian Bioscience®) and associated qPCR cycler. Alternatively, a Qubit Fluorometer or a similar device can be used, but concentrations should be compared to qPCR in the first experiments to establish a normalization factor for further experiments, if needed.

3.6. Sequencing

- Reagents for library denaturation and sequencing: Please order directly from Illumina, Inc. We recommend using the MiSeq Reagent kit v3 (600-cycle), MiSeq Reagent kit v2 (500-cycle) or MiSeq Reagent kit Nano v2 (500-cycle) for best performance.
- Optional but recommended: PhiX Control v3 spiked in at 15-20% (Illumina).

4. Laboratory protocol

4.1. First PCR – Amplification of target region

4.1.1. Reaction setup

Dilute the DNA samples to a concentration of 0.5-5ng/μl and set up the PCR reaction as shown in Table 1.

The MIB1 - V3V4 Mix contains different heterogeneity spacers ensuring sequence heterogeneity in the library pool which is necessary for Illumina sequencing.

Note: Set up PCR reaction on ice.

Table 1 - Reaction Setup PCR1

Component	Volume for 1 reaction [μl]	Volume for 10 reactions incl. 5% excess [μl]
16S Master Mix 2x	25.0	262.5
MIB1 - V3V4 Mix	3.2	33.6
DNA template (0.5-5 ng/μl)	5.0	-
Nuclease-free water	16.8	176.4
TOTAL VOLUME	50.0	472.5 (use 45 μl/reaction, add 5μl DNA template)

4.1.2. PCR program

Create a PCR program according to Table 2. Use a heated lid (>100°C); if applicable, set the heating ramp rate to max. 2.5°C and the cooling ramp rate to max. 1.5°C.

Table 2 - PCR Program PCR1

Temperature	Time	Comment
95°C	03:00	Initial Denaturation
95°C	00:15	20 cycles
55°C	00:15	
72°C	00:30	
72°C	10:00	Final elongation
4°C	Hold	

4.2. Cleanup of PCR1 products

4.2.1. Preparation

- Let the Magnetic Beads equilibrate at room temperature by transferring the bottle to the bench at least 30 min prior to use.
- Always prepare fresh 80% ethanol.
- Thoroughly resuspend the beads immediately prior to use.

4.2.2. Cleanup procedure

1. Add 60µl Magnetic Beads directly to PCR1 products (50µl) and mix well by pipetting up and down at least 10 times.

Note: *If using less than 50µl PCR1 product adjust the bead volume accordingly: [bead] : [sample] ratio = 1.2 : 1.*

2. Incubate for 5 minutes at room temperature.
3. Place the tubes/plate on a magnetic stand to immobilize the beads. Incubate at room temperature until the solution is completely clear (approximately 1-2 minutes).
4. Aspirate and discard the clear supernatant without touching the beads.
5. Keep the tubes/plate on the magnetic stand and add 150µl 80% ethanol to each tube/well. Do not resuspend the beads pellet.
6. Incubate at room temperature for 30 seconds on the magnetic stand.
7. Aspirate and discard the supernatant without touching the beads.
8. Repeat Steps 6 to 8 one more time for a total of 2 washes.
9. Leave the tubes on the magnetic stand, remove residual ethanol with a pipette and dry the beads for approximately 3 minutes.

Important: *Residual ethanol may interfere with downstream applications. Do not over-dry the beads as this will decrease yield. The bead pellet is dry when the appearance of the surface changes from shiny to matte.*

10. Remove the tubes/plate from the magnetic stand.
11. Elute the samples in 25µl nuclease-free water. Mix well by pipetting up and down 10 times.
12. Incubate at room temperature for 3 minutes.
13. Place the tubes on a magnetic stand and incubate at room temperature until the beads are completely cleared from solution (approximately 1-2 minutes).
14. Transfer 21.8µl of the cleared supernatant to a new PCR tube.

4.2.3. (Optional) Check PCR1 product size (e.g. on a Fragment Analyzer)

The expected size of the post-PCR1 library is approximately 550 bp (See Figure 2).




Figure 2 – Fragment Analyzer profile of PCR1 product.

4.3. Second PCR – Indexing

4.3.1. Reaction setup

PCR1 product is used as template for PCR2. Set up the PCR reaction as stated in Table 3. For each sample, select a unique combination of indexing forward and reverse primers. The variation of PCR2 forward and reverse primers enables for pooling 96 samples. If a higher number of samples is intended to be pooled, please order kits with different Indexing Primer Sets.

Note: Set up PCR reaction on ice.

Important: Make sure that all individual libraries sequenced in the same pool have a unique forward / reverse indexing primer combination and make note of the sample – dual-index combination (see Appendix) for the demultiplexing of the individual samples after sequencing.

If only a small number of samples is to be sequenced, please use combinations as described in the Appendix for optimal balancing.

Table 3 - Reaction Setup PCR2

Component	Volume for 1 reaction [µl]
ViennaLab 16S Master Mix 2x	25.0
Indexing Forward primer: MIB2(A/B/C)-Fx	1.6
Indexing Reverse primer: MIB2X-Rx	1.6
Product of PCR1	21.8
TOTAL VOLUME	50.0

4.3.2. PCR program

Create a PCR program according to Table 2. Use a heated lid (>100°C); if applicable, set the heating ramp rate to max. 2.5°C and the cooling ramp rate to max. 1.5°C.

Table 4 - PCR Program PCR2

Temperature	Time	Comment
95°C	03:00	Initial Denaturation
95°C	00:15	10 cycles
55°C	00:15	
72°C	00:30	
72°C	10:00	Final elongation
4°C	hold	

4.4. Cleanup of PCR2 products

4.4.1. Preparation

- Let the Magnetic Beads equilibrate at room temperature by transferring the bottle to the bench at least 30 min prior to use.
- Always prepare fresh 80% ethanol.
- Thoroughly resuspend the beads immediately prior to use.

4.4.2. Cleanup procedure

1. Add 60µl Magnetic Beads directly to PCR1 products (50µl) and mix well by pipetting up and down at least 10 times.

Note: *If using less than 25µl PCR2 product adjust bead volume accordingly:
[bead] : [sample] ratio = 1.2 : 1.*

2. Incubate for 5 minutes at room temperature.
3. Place the tubes/plate on a magnetic stand to immobilize the beads. Incubate at room temperature until the solution is completely clear (approximately 1-2 minutes).
4. Aspirate and discard the clear supernatant without touching the beads.
15. Keep the tubes/plate on the magnetic stand and add 150µl 80% ethanol to each tube/well. Do not resuspend the beads pellet.
5. Incubate at room temperature for 30 seconds on the magnetic stand.
6. Aspirate and discard the supernatant without touching the beads.
7. Repeat Steps 6 to 8 one more time for a total of 2 washes.
8. Leave the tubes on the magnetic stand, remove residual ethanol with a pipette and dry the beads for approximately 3 minutes.

Important: *Residual ethanol may interfere with downstream applications. Do not over-dry the beads as this will decrease yield. The bead pellet is dry when the appearance of the surface changes from shiny to matte.*

9. Remove the tubes/plate from the magnetic stand.
10. Elute the samples in 25µl nuclease-free water. Mix well by pipetting up and down 10 times.
11. Incubate at room temperature for 3 minutes.
12. Place the tubes on a magnetic stand and incubate at room temperature until the beads are completely cleared from solution (approximately 1-2 minutes).
13. Transfer 22µl of the cleared supernatant containing cleaned-up PCR products to a new tube.

4.4.3. (Optional) Check PCR2 product size (e.g. on a Fragment Analyzer)

The expected size of the post-PCR2 library is approximately 600 bp (See Figure 3).




Figure 3 - Fragment Analyzer profile of PCR2 product.

4.5. Library quantification and pooling

4.5.1. Library quantification

Library quantification by qPCR is recommended, e.g. with JetSeq™ Library Quantification Lo-ROX Kit (meridian Bioscience®) and associated qPCR cycler. Alternatively, a Qubit Fluorometer or a similar device can be used, but concentrations should be compared to qPCR in the first experiments to establish a normalization factor for further experiments, if needed.

4.5.2. Library dilution and pooling

Dilute all individual libraries to 4 nM and pool for sequencing equimolarly or according to the required amount of data output.

For the assessment of the number of samples that can be sequenced with the used sequencing chemistry, the calculation of the sample concentration and pooling, and the assistance with the sample sheet generation, please refer to the ViennaLab MicrobeCalc™ Rev1.4 Excel file that can be downloaded at <https://viennalab.com/support/ngs-assays>.

Important: Use ViennaLab MicrobeCalc™ Rev1.4 (or later) for kits using IFU Rev3.0, and ViennaLab MicrobeCalc™ with kits using IFU Rev2.0.

5. Sample Sheet and sequencing recommendations

- For library pool denaturation and loading refer to Illumina System Guides.
- A spike-in of 15-20% PhiX is recommended.
- For sample sheet preparation see index sequences listed in Table 5 or Table 6 depending on your sequencing instrument. For MiSeq the ViennaLab MicrobeCalc™ Excel file can be used (see 4.5.2).



Table 5 - Index sequences for MiSeq, MiniSeq (Rapid Reagent kits), HiSeq 2000/2500 or NovaSeq 6000 (v1.0 reagent kits)

i5 Primer	i5 index sequence	i7 Primer	i7 index sequence
MIB2A-F1	TAAGACAC	MIB2X-R1	ACGCTACT
MIB2A-F2	CTGCGTGT	MIB2X-R2	CGAGCGAC
MIB2A-F3	TGTTCTCT	MIB2X-R3	GTCTATGA
MIB2A-F4	CTAATCGA	MIB2X-R4	TATAGCGA
MIB2A-F5	CGTACTA	MIB2X-R5	CGAGAGTT
MIB2A-F6	ACTATCTG	MIB2X-R6	GACATAGT
MIB2A-F7	TGAACCTT	MIB2X-R7	ACTCACTG
MIB2A-F8	TAAGTTCC	MIB2X-R8	TAGTCTCC
MIB2B-F9	TACGAGAC	MIB2X-R9	TGAGTACG
MIB2B-F10	CGTGAGTG	MIB2X-R10	CTGCGTAG
MIB2B-F11	GACACCGT	MIB2X-R11	ACTACGAC
MIB2B-F12	CTACTATA	MIB2X-R12	GTCTGCTA
MIB2B-F13	GATCGTGT		
MIB2B-F14	GTCAGATA		
MIB2B-F15	AGAGTCAC		
MIB2B-F16	TCATCGAG		
MIB2C-F17	ATCGTACG		
MIB2C-F18	TCGACGAG		
MIB2C-F19	GGATATCT		
MIB2C-F20	TAGACCTA		
MIB2C-F21	ACGTCTCG		
MIB2C-F22	TAGCGAGT		
MIB2C-F23	CTAGAACA		
MIB2C-F24	TGCTAAGT		

Table 6 - Index sequences for iSeq100, MiniSeq (Standard reagent kits), NextSeq, HiSeq 3000/4000/X or NovaSeq 6000 (v1.5 reagent kits)

i5 Primer	i5 index sequence	i7 Primer	i7 index sequence
MIB2A-F1	GTGTCTTA	MIB2X-R1	ACGCTACT
MIB2A-F2	ACACGCAG	MIB2X-R2	CGAGCGAC
MIB2A-F3	AGAGAACA	MIB2X-R3	GTCTATGA
MIB2A-F4	TCGATTAG	MIB2X-R4	TATAGCGA
MIB2A-F5	TAGTAACG	MIB2X-R5	CGAGAGTT
MIB2A-F6	CAGATAGT	MIB2X-R6	GACATAGT
MIB2A-F7	AAGGTTCA	MIB2X-R7	ACTCACTG
MIB2A-F8	GGAACCTA	MIB2X-R8	TAGTCTCC
MIB2B-F9	GTCTCGTA	MIB2X-R9	TGAGTACG
MIB2B-F10	CACTCACG	MIB2X-R10	CTGCGTAG
MIB2B-F11	ACGGTGTC	MIB2X-R11	ACTACGAC
MIB2B-F12	TATAGTAG	MIB2X-R12	GTCTGCTA
MIB2B-F13	ACACGATC		
MIB2B-F14	TATCTGAC		
MIB2B-F15	GTGACTCT		
MIB2B-F16	CTCGATGA		
MIB2C-F17	CGTACGAT		
MIB2C-F18	CTCGTCGA		
MIB2C-F19	AGATATCC		
MIB2C-F20	TAGGTCTA		
MIB2C-F21	CGAGACGT		
MIB2C-F22	ACTCGCTA		
MIB2C-F23	TGTTCTAG		
MIB2C-F24	ACTTAGCA		

6. Quick Guide to sequencing data analysis

- 6.1. a) Upon first use please navigate to <https://microbiome.viennalab.com> and register for the ViennaLab Microbiome Analysis Webtool providing a valid email address and the Activation code you find on the inner side of the kit box lid.
b) If you are an already registered user, please enter the Activation code provided with the kit in your Account profile.
- 6.2. Upload the raw sequencing data files. Supported file formats include *.fastq and *.fastq.gz. If you have paired input files, check the box “Paired input files” and click on  .
- 6.3. You will then be asked to confirm the deduction of 1 sample credit from your balance for each single or two paired sequencing data files by clicking on  .
***Important:** Do not close the browser window or redirect or reload the page while files are being uploaded, as this would interrupt the upload procedure.*
- 6.4. When upload and processing are finished sample status will turn from “Processing” to “Ready”.
- 6.5. You can now review your results and generate reports.

For technical support please contact ViennaLab through the local distributor or directly at techhelp@viennalab.com.

APPENDIX

Tables for indexing primer combinations

- If preparing only a small number of samples (e.g. less than 16 samples), use index combinations with consecutive numbers given in the tables below (e.g. use primers F1-F4 and R1-R2 for samples 1-8 as indicated, and F1-F4 and R3-R4 for samples 9-16 and so on for kit [9-131]).
- Mark which index combinations have already been used and use each primer combo only once.

[9-131]

MIB2-	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	R11	R12
F1	1	2	9	10	65	66	73	74	33	34	41	42
F2	3	4	11	12	67	68	75	76	35	36	43	44
F3	5	6	13	14	69	70	77	78	37	38	45	46
F4	7	8	15	16	71	72	79	80	39	40	47	48
F5	49	50	57	58	17	18	25	26	81	82	89	90
F6	51	52	59	60	19	20	27	28	83	84	91	92
F7	53	54	61	62	21	22	29	30	85	86	93	94
F8	55	56	63	64	23	24	31	32	87	88	95	96

[9-132]

MIB2-	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	R11	R12
F9	1	2	9	10	65	66	73	74	33	34	41	42
F10	3	4	11	12	67	68	75	76	35	36	43	44
F11	5	6	13	14	69	70	77	78	37	38	45	46
F12	7	8	15	16	71	72	79	80	39	40	47	48
F13	49	50	57	58	17	18	25	26	81	82	89	90
F14	51	52	59	60	19	20	27	28	83	84	91	92
F15	53	54	61	62	21	22	29	30	85	86	93	94
F16	55	56	63	64	23	24	31	32	87	88	95	96

[9-133]

MIB2-	R1	R2	R3	R4	R5	R6	R7	R8	R9	R10	R11	R12
F17	1	2	9	10	65	66	73	74	33	34	41	42
F18	3	4	11	12	67	68	75	76	35	36	43	44
F19	5	6	13	14	69	70	77	78	37	38	45	46
F20	7	8	15	16	71	72	79	80	39	40	47	48
F21	49	50	57	58	17	18	25	26	81	82	89	90
F22	51	52	59	60	19	20	27	28	83	84	91	92
F23	53	54	61	62	21	22	29	30	85	86	93	94
F24	55	56	63	64	23	24	31	32	87	88	95	96

Version	Amendments
rev 3.0	<ul style="list-style-type: none">- Update of sample input amount- Change of PCR1 primers from individual primer tubes to one primer pool- Change of PCR1 cycle number- Change of beads clean up ratio and volumes after PCR1 and PCR2- Change of sample input volume into PCR2- Change of PCR2 primer sequences- Change of ViennaLab MicrobeCalc™ tool to Rev1.4 for demultiplexing

Notes

Notes

Notes

REF



9-131	16S Microbiome NGS Assay [Set A]	96 reactions
9-132	16S Microbiome NGS Assay [Set B]	96 reactions
9-133	16S Microbiome NGS Assay [Set C]	96 reactions

Distributed by:



Manufacturer

ViennaLab Diagnostics GmbH

Gaudenzdorfer Guertel 43-45

A-1120 Vienna, Austria

t: (+43-1) 8120156-0

e: info@viennalab.com



www.viennalab.com