

DECLARATION OF CONFORMITY



BIOMEDICAL SYSTEMS

370 West 1700 South
Logan, Utah 84321 USA
435 752 6011
800 453 2725
fax 435 752 4127
web www.wescor.com

Manufacturer's Name: Wescor, Inc
Manufacturer's Address: 370 W 1700 S
Logan, Utah 84321-5294
USA

European Authorized Representative: **MT Promedt Consulting, GmbH**
Altenhofstr. 80
D-66386 St. Ingbert / Germany
Tel.: +49 6894 - 58 10 20
Fax: +49 6894 - 58 10 21
www.mt-procons.com

Declares the product

Product Name: Vapro[®] Vapro Pressure Osmometer
Model Number: 5600
Type of Equipment: AC line powered Osmometer

Conforms to the requirements of the following Standards and Specifications:

IEC 61010-1 (2001)
IEC 61010-2-081 (2001)
IEC 61010-2-101 (2002)

**Thereby complying with the In-vitro Diagnostic Directive 98/79/EC
and therefore bearing the IVD / CE Marking.**

Accessories and Supplies

57-0006-01	Vapro Applications Manual
AC-011	9/64-inch Hex Driver
AC-036	Forceps
AC-037	10 uL Pipettor
AC-061	Ampule Organizer
AC-062	Sample Holder, 7 mm X 1.25 mm deep, models 5520, 5500, 5600

AC-063 Sample Holder, for 2 L Samples, 4.25 mm dia. X 1.12 mm deep, models 5520, 5500, 5600
AC-064 Sample Holder, 7 mm dia. X 2.5 mm deep, models 5520, 5500, 5600
AC-065 Sample Holder, 9.5 mm dia. X 4.5 mm deep, models 5520, 5500, 5600
AC-174 TC Head Assembly 0-3200 mmol/kg, for Vapro 5600
AC-176 Lab Report Software, Vapro
DOC-00588 Instructions For Use
OA-010 100 mmol/kg Opti-Mole Osmolality Ampule Standard
OA-029 290 mmol/kg Opti-Mole Osmolality Ampule Standard
OA-100 1000 mmol/kg Opti-Mole Osmolality Ampule Standard
SS-003 Cleaning Solution
SS-006 De-ionized Water
SS-026 Blow Clean
SS-033 Paper Sample Discs
SS-036 Micropipettor Disposable Tips
SS-223 Thermocouple Cleaning Solution
SS-238 Desiccant Filter Cartridge for Vapro 5600
SS-239 Waste Reservoir for Vapro 5600, Colloid 4500
SS-240 Supply Reservoir for Vapro 5600, Colloid 4500



J R Petersen, QA \ RA Manager
Wescor, Inc
370 West 1700 South
Logan, UT 84321

4.23.010
Date