




## • Certificate of Analysis

Product Name : IVD Matrix HCCA-portioned  
Part No. 8290200  
LOT No.: 6020222008  
Use by<sup>1)</sup>: 2024-01-31  
Date of Production: 2022-02-01  
Storage Conditions : Store in a refrigerator at 2°C to 8°C on arrival.

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "Escherichia coli" as best match with a log score value $\geq 2.000$ .	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value $\leq 1.699$ .	Passed

**Remarks:** <sup>1)</sup> IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG, Germany  
Fahrenheitstr. 4 – 28359 Bremen phone  
+49-421-2205-0



Dr. M. Kostrzewa, SAFETY OFFICER

QC release date: 2022-02-15

Document Revision: 3017659 B - DCR-03726