

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert® MRSA NXG	a		
Cepheid Catalogue Part No.: GXMRSA-NXG-CE-10			
Kit Lot No.: 1000169762			
Cartridge Lot No.: 05206			
Kit Expiration Date: 2021-08-08			
Legal Manufacturer	Manufacturing Facility	○ Solna	Sunnyvale
Cepheid 904 Caribbean Drive	Cepheid 904 Caribbean Drive		

Functional Testing

Sunnyvale, CA 94089 USA

Test Description	Acceptance Criteria	Test Result
LOW Positive	MRSA detected	Passed
HIGH Positive	MRSA detected	Passed
Negative	MRSA not detected	Passed

Sunnyvale, CA 94089 USA

Equants	10/15	119
Signature of Quality Assurance		Date
Name: Emaela S. Quan-Tong		
Title: Quality Systems Specialis	st	