

Name: Samaneh Vahid

Title: QA Analyst

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Directive and the Canadian Medical Device Regulations (CMDR).				
Product Name: Xpert® Norovirus				
Cepheid Catalogue Part No.: GXNOV-CE-10				
Kit Lot No.: 1000	172515			
Cartridge Lot No.: 152	202			
Kit Expiration Date: 2	2020-11-22			
Legal Manufacturer		Manufacturing Facility	Solna (Sunnyvale
Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden		
Functional Testing				
Test Description	Acceptance Criteria			Test Result
LOW Positive	NORO GI DETECTED;NORO GII DETECTED			Passed
HIGH Positive	NORO GI DETECTED;NORO GII DETECTED			Passed
Negative	NORO GI NOT DETECTED;NORO GII NOT DETECTED			Passed
11 0.	20101017			
Signature of Quality Assurance Date				