

## • Certificate of Analysis

Product Name : IVD Bacterial Test Standard

Part No. 8290190

LOT No.: 7030320012

Use by<sup>1)</sup>: 2021-10-31

Date of Production: 2020-11-12

QC release date: 2020-11-16

Storage Conditions : Store at -18 °C or below immediately on arrival.

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Verification of the calibration points by MALDI-TOF MS	Within a mass range of $\pm 900$ ppm	Passed
Relative intensity of reference proteins No. 3 by MALDI-TOF MS	$0.20 \pm 0.05$	Passed
Relative intensity of reference proteins No. 4 by MALDI-TOF MS	$0.22 \pm 0.05$	Passed
MALDI Biotyper identification	Classification Result Report lists "Escherichia coli" as best match.	Passed
	log(score) Value $\geq 2.2$ of the best match	Passed

### Remarks:

- 1) IVD BTS is shipped at ambient temperature but must be stored at  $-18^{\circ}\text{C}$  or below immediately on arrival. The expiry date on the package is valid for the enclosed IVD BTS when stored at  $-18^{\circ}\text{C}$  or below on arrival. Frozen, dissolved Bruker IVD BTS can be stored for up to 5 months at  $-18^{\circ}\text{C}$  or below. See instruction for use at [www.bruker.com/care](http://www.bruker.com/care).

Bruker Daltonik GmbH, Germany  
Fahrenheitstr. 4 – 28359 Bremen  
phone +49-421-2205-0

  
Bruker Daltonik GmbH  
Fahrenheitstraße 4  
D-28359 Bremen  
Tel: +49 (0) 421 - 22 05 - 0  
Dr. Markus Kostrzewa, SAFETY OFFICER