

# • Certificate of Analysis

Product Name : **IVD Matrix HCCA-portioned**

Part No. **8290200**  
 LOT No.: **0000275020**  
 Use by<sup>1)</sup>: **2019-05**  
 Date of Production: **2017-06-30**  
 Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
<b>1. Amount of HCCA content</b>	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
<b>2. MALDI Biotyper identification of a 1:20 diluted IVD BTS</b>	Classification Result Report lists " <i>Escherichia coli</i> " as best match with a log score value $\geq 2.000$ .	Passed
<b>3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper</b>	Classification Result Report lists "not reliable identification." with a log score value $\leq 1.699$ .	Passed

**Remarks:** <sup>1)</sup> IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0

**Bruker Daltonik GmbH**  
 Fahrenheitstraße 4,  
 D-28359 Bremen  
 Tel: +49 (0) 421 - 22-05-0

Dr.M. Kostrzewa, SAFETY OFFICER

QC release date: 2017-08-01