

# • Certificate of Analysis

Product Name : IVD Matrix HCCA-portioned

Part No. 8290200  
 LOT No.: 0000287504  
 Use by<sup>1)</sup>: 2020-08  
 Date of Production: 2018-09-18  
 Storage Conditions : Store in a refrigerator at 2°C to 8°C on arrival.

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "Escherichia coli" as best match with a log score value $\geq 2.000$ .	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value $\leq 1.699$ .	Passed

**Remarks:** <sup>1)</sup> IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0

  
 Bruker Daltonik GmbH  
 Fahrenheitstraße 4  
 D-28359 Bremen  
 Tel: +49 (0) 421 - 22 05 - 0  
 Dr. M. Kostrzewa, SAFETY OFFICER

QC release date: 2018-10-08