

• Certificate of Analysis

Product Name : **IVD Matrix HCCA-portioned**

Part No. **8290200**

LOT No.: **0000287520**

Use by¹⁾: **2020-11-30**

Date of Production: **2018-12-12**

Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

| <u>Test:</u> | <u>Specification</u> | <u>LOT Results</u> |
|---|---|--------------------|
| 1. Amount of HCCA content | AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1 | Passed |
| 2. MALDI Biotyper identification of a 1:20 diluted IVD BTS | Classification Result Report lists "Escherichia coli" as best match with a log score value ≥ 2.000 . | Passed |
| 3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper | Classification Result Report lists "not reliable identification." with a log score value ≤ 1.699 . | Passed |

Remarks: ¹⁾ IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany
Fahrenheitstr. 4 – 28359 Bremen
phone +49-421-2205-0


Bruker Daltonik GmbH
 Fahrenheitstraße 4
 D-28359 Bremen
 Dr. M. Kostzowa, SAFETY OFFICER

QC release date: 2019-01-18