

• Certificate of Analysis

Product Name : **IVD Matrix HCCA-portioned**

Part No. **8290200**

LOT No.: **6020221031**

Use by¹⁾: **2023-08-31**

Date of Production: **2021-09-02**

Storage Conditions : **Store in a refrigerator at 2°C to 8°C on arrival.**

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "Escherichia coli" as best match with a log score value ≥ 2.000 .	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value ≤ 1.699 .	Passed

Remarks: ¹⁾ IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG, Germany
Fahrenheitstr. 4 – 28359 Bremen phone
+49-421-2205-0



• Bruker Daltonics GmbH & Co. KG
Fahrenheitstr. 4
D-28359 Bremen
Phone + 49 (0)421- 2205-0
Fax +49 (0)421-2205-377

Dr. M. Kostrzewa, SAFETY OFFICER

QC release date: 2021-09-07