

# • Certificate of Analysis

Product Name : IVD Matrix HCCA-portioned  
 Part No. 8290200  
 LOT No.: 6020221051  
 Use by<sup>1)</sup>: 2023-09-30  
 Date of Production: 2021-10-20  
 Storage Conditions : Store in a refrigerator at 2°C to 8°C on arrival.

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "Escherichia coli" as best match with a log score value $\geq 2.000$ .	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value $\leq 1.699$ .	Passed

**Remarks:** <sup>1)</sup> IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonics GmbH & Co. KG, Germany  
 Fahrenheitstr. 4 – 28359 Bremen phone  
 +49-421-2205-0



• Bruker Daltonics GmbH & Co. KG  
 Fahrenheitstr. 4  
 D-28359 Bremen  
 Phone + 49 (0)421 2205-0  
 Fax +49 (0)421 2205-370

Dr. M. Kostrzewa SAFETY OFFICER

QC release date: 2021-10-21