

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Directive and the Canadian Medical Device Regulation	is (CMDR).		
Product Name:_ Xpert® SA Nasal Complete			
Cepheid Catalogue Part No.: GXSACOMP-CE-10			
Kit Lot No.: 1000165931			
Cartridge Lot No.: 22111			
Kit Expiration Date: 2021-08-01			
Legal Manufacturer	Manufacturing Facility	○ Solna	Sunnyvale
Cepheid AB Röntgenvägen 5 SE-171 54 Solna	Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA		a .
Sweden			

Functional Testing

Test Description	Acceptance Criteria	Test Result
Positive	MRSA Positive; SA Positive	Passed
Negative	MRSA Negative; SA Negative	Passed

Signature of Quality-Assurance	8/23/	/2019 Date
Name: ThuyTien Nguyen		
Title: Quality Systems Specialist		