

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

| Product Name: Xpert® | SA Nasal Complete | | |
|---|----------------------------|---|-------------|
| Cepheid Catalogue Part | No.: GXSACOMP-CE- | 10 | |
| Kit Lot No.: 100024 | 2183 | | |
| Cartridge Lot No. 2340 | 6 | | |
| Kit Expiration Date: 20 | 22-12-25 | | |
| Legal Manufacturer | | Manufacturing Facility () Solna (| Sunnyvale |
| Cepheid AB Röntgenvägen 5 SE-171 54 Solna | | Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA | · |
| Sweden | | | |
| Functional Testing | | | |
| Test Description | Acceptance Criteria | | Test Result |
| Positive | MRSA Positive; SA Positive | | Passed |
| Negative | MRSA Negative; SA Negative | | Passed |

Signature of Quality Assurance

Name: Molly Doan

Title: Quality System Specialist III