


CERTIFICATE OF ANALYSIS

Product Name: Microbank™ (With 2D Barcodes) – Mixed		Lot Number: C20035M
Product Code: PL.170C/M		Expiration Date: 2023-08-31
Product Code	Quantity	Components
PL.170C/M	80	Vials printed with unique 2D barcodes containing approximately 25 beads and 1 ml of cryopreservative
	1	Numbered storage container with a unique 2D barcode
Additional Information		
Storage Conditions:	15-30°C	
Precautions:	Refer to Instructions for Use	
Quality Control Testing		
	Criteria	Result
Sterility*	Conforms	PASS
Visual Inspection	Conforms	PASS
Barcode Verification/ Barcode Sequencing	Conforms	PASS
	Recovery of Organism	Result
Organism		
<i>Escherichia coli</i> ATCC 25922	Satisfactory	PASS
<i>Staphylococcus aureus</i> ATCC 25923	Satisfactory	PASS
<i>Klebsiella pneumoniae</i> ATCC 13883	Satisfactory	PASS
<i>Clostridium perfringens</i> ATCC 13124	Satisfactory	PASS
<i>Neisseria gonorrhoeae</i> ATCC 43069	Satisfactory	PASS
*This product has been sterilized by GAMMA irradiation in a validated sterilization cycle and meets the established validated specified dose.		
Release		
2021-12-22		
Date	Quality Control	

Canada ☎ 800 268 2341 Fax 905 731 0206
 20 Mural Street, Unit #4, Richmond Hill, Ontario,
 L4B 1K3
 support@pro-lab.com

U.S.A. ☎ (512) 832 9145 Fax (512) 832 6424
 21 Cypress Blvd., Suite 1155, Round Rock,
 Texas 78665-1046
 support@pro-lab.us

U.K. ☎ 0151 353 1613 Fax 0151 353 1614
 3 Bassendale Road, Bromborough, Wirral,
 Merseyside CH62 3QL
 uksupport@pro-lab.com