



# CERTIFICATE OF ANALYSIS

MIC\_COA\_SI390900#SCR14106.pdf

Product Code <b>SI 390.900</b>	Name <b>URO-QUICK SCREENING KIT 360 TEST</b>	Quantity manufactured <b>27</b>	
Lot Number <b>SCR14106</b>	Manufacturing Date <b>17/04/2014</b>	Expiry Date <b>17/06/2015</b>	CAP color <b>Green</b>
			Printed Band <b>White</b>

## Eugonic Broth

Lot Number **SCR14106**  
 Expiry Date **17/06/2015**  
 CAP color **Green**  
 Printed Band **White**

	Physical parameters <i>(Volume, transparency, colour )</i>	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% ≤ 35%	CV% ≤ 35%
Found	OK	OK	32.4%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Date of Control **12/05/2014**

Date of Approval **13/05/2014**

Tested **Cirio Alberto**

Quality Control Manager **Cossettini Marco**

**ALIFAX S.p.A.**  
 via Petrarca 2/1 - 35020 POLVERARA (PD) - ITALY  
 Tel. +39 0490992000  
 Web site: [www.alifax.com](http://www.alifax.com) e-mail: [info@alifax.com](mailto:info@alifax.com)



**SIRE Analytical Systems S.r.l.**  
 via Merano 30 - 33045 Nimis (UDINE) - ITALY  
 Tel. +39 0432547454