



# CERTIFICATE OF ANALYSIS

MIC\_COA\_SI390900#SCR16014.pdf

Product Code <b>SI 390.900</b>	Name <b>KIT URO QUICK SCREENING 360 TEST</b>	Quantity manufactured <b>24</b>	
-----------------------------------	---	------------------------------------	--

Lot Number <b>SCR16014</b>	Manufacturing Date <b>02/02/2016</b>	Expiry Date <b>29/03/2017</b>	CAP color <b>Green</b>	Printed Band <b>White</b>
-------------------------------	---	----------------------------------	---------------------------	------------------------------

## Eugonic Broth

Lot Number **SCR16014**  
Expiry Date **29/03/2017**  
CAP color **Green**  
Printed Band **White**

	Physical parameters <i>(Volume, transparency, colour )</i>	Decontamination at 37°C after 72 hours	Growth inter batch	Growth intra batch
Expected	Compliant	No growth	CV% <= 35%	CV% <= 35%
Found	OK	OK	5.4%	< 5%

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Date of Control **08/02/2016**

Date of Approvation **08/02/2016**

Tested **Cirio Alberto**

Quality Control Manager **Ferino Rudy**

### Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy  
Tel (+39) 049 0992000 - Fax (+39) 049 5855434  
Mail PEC: info.alifax@pec.alifax.com  
[info@alifax.com](mailto:info@alifax.com) - [www.alifax.com](http://www.alifax.com)



### Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy  
Tel (+39) 0432 547454 - Fax (+39) 0432 547378  
Mail PEC: info.udine@pec.alifax.com  
[info.udine@alifax.com](mailto:info.udine@alifax.com)