



CERTIFICATE OF ANALYSIS

MIC_COA_SI390900#SCR16100.pdf

| | | | |
|-----------------------------------|---|------------------------------------|--|
| Product Code SI 390.900 | Name KIT URO QUICK SCREENING 360 TEST | Quantity manufactured 27 | |
|-----------------------------------|---|------------------------------------|--|

| | | | | |
|-------------------------------|---|----------------------------------|---------------------------|------------------------------|
| Lot Number SCR16100 | Manufacturing Date 06/06/2016 | Expiry Date 05/08/2017 | CAP color Green | Printed Band White |
|-------------------------------|---|----------------------------------|---------------------------|------------------------------|

Eugonic Broth

Lot Number **SCR16100**
Expiry Date **05/08/2017**
CAP color **Green**
Printed Band **White**

| | Physical parameters (Volume, transparency, colour) | Decontamination at 37°C after 72 hours | Growth inter batch | Growth intra batch |
|----------|---|---|--------------------|--------------------|
| Expected | Compliant | No growth | CV% <= 35% | CV% <= 35% |
| Found | OK | OK | 14.1% | 7.4% |

Check of the presence of the magnetic stirrer in each vial contained in the kit: **YES**

Storage Conditions: **min 4°C - Max 30°C**

Compliance: **YES**

Date of Control **16/06/2016**

Date of Approval **16/06/2016**

Controlled By **Cirio Alberto**

Approved By **Ferino Rudy**

Registered Office

Via Petrarca, 2/1 - 35020 Polverara (PD) Italy
Tel (+39) 049 0992000 - Fax (+39) 049 5855434
Mail PEC: info.alifax@pec.alifax.com
info@alifax.com - www.alifax.com



Manufacturing Site

Via Merano, 30 - 33045 Nimis (UD) Italy
Tel (+39) 0432 547454 - Fax (+39) 0432 547378
Mail PEC: info.udine@pec.alifax.com
info.udine@alifax.com